



Test Report

No.: SZHL260401809501OT

Date: MAY 09, 2026

Page 1 of 3

Dongguan jiajie medical tech Co.,Ltd

Floor 3, Building 1, San Fengyuan Park,NO.3, Datang Hongfa Street, , Da Lingshan Town, Dongguan, Guangdong, China

The following samples were submitted and identified by/on behalf of the client as:

Sample Descriptions : Silicone tube
Country of Origin : China
Country of Destination : Bangladesh
Request Age Grading : 30
Importer : SI TRADE & Engineering
Supplier : Dongguan jiajie medical tech Co.,Ltd
Manufacturer : Dongguan jiajie medical tech Co.,Ltd

SGS Order No : SZHL26040180950T
Sample Receiving Date : APR 28, 2026
Testing Period : APR 28, 2026 TO MAY 09, 2026

Test Requested		Conclusion
1	US FDA 21 CFR 177.2600 - Total Extractives	PASS

***** FOR FURTHER DETAILS, PLEASE REFER TO THE FOLLOWING PAGE(S) *****

Signed for and on behalf of
SGS-CSTC Standards Technical Services Co., Ltd. Shenzhen Branch

Wang Juan, Zoe
Senior Engineer



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone:(86-755) 8307 1443, or email: CN.Doccheck@sgs.com

SGS-CSTC Standards Technical Services Co., Ltd.
Shenzhen Branch Testing Center Hardlines

Room 101-501, Plant 4 & Room 101, Plant 2 & Room 101, Plant 3 & Room 301-501, Plant 3, Jianghao (Bantian) Industrial Plant Area, No.430, Jihua Road, Bantian Community, Bantian Street, Longgang District, Shenzhen, Guangdong, China 518129 www.sgs.com.cn
中国·广东·深圳市龙岗区坂田街道坂田社区吉华路 430 号江灏 (坂田) 工业厂区厂房 4 号 101-901、2 号 101、3 号 101、3 号 301-501 邮编: 518129 t:(86-755) 253 sgs.china@sgs.com

Member of the SGS Group (SGS SA)

Test Result(s):

Test Part Description:

SN ID	Sample No.	Description	Material (claimed by the client)
SN1	A1	Transparent silicone (Tube)	Silicone Rubber

Remarks:

- (1) mg/dm² = milligram per square decimeter
 mg/inch² = milligram per square inches
 µg/cm² = microgram per square centimeter
 mg/L = milligram per litre
 mg/kg = milligram per kilogram
- (2) °C = degree Celsius
 °F = degree Fahrenheit
- (3) <= less than
- (4) RL = Reporting Limit
- (5) ND = Not Detected (< RL).

US FDA 21 CFR 177.2600 - Total Extractives

Test Method: With reference to US FDA 21 CFR 177.2600.

Simulant used	Test Condition	Limit	Unit(s)	RL	A1	Comment
Distilled water	Reflux Temperature 7.0 hrs	20	mg/inch ²	0.5	ND	Pass
Distilled water Succeeding extraction	Reflux Temperature 2.0 hrs	1	mg/inch ²	0.5	ND	Pass



SGS-CSTC Shenzhen Technical Services Co., Ltd.
 Shenzhen Branch Testing Center Hardlines

Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com/en/Terms-and-Conditions/Terms-e-Documents.aspx>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 30 days only.

Attention: To check the authenticity of testing / inspection report & certificate, please contact us at telephone:(86-755) 8307 1443, or email: CN.Doccheck@sgs.com

Room 101-501, Plant 4 & Room 101, Plant 2 & Room 101, Plant 3 & Room 301-501, Plant 3, Jianghao (Bantian) Industrial Plant Area, No.430, Jihua Road, Bantian Community, Bantian Street, Longgang District, Shenzhen, Guangdong, China 518129 www.sgs.com
 中国·广东·深圳市龙岗区坂田街道坂田社区吉华路 430 号江灏 (坂田) 工业厂区厂房 4 号 101-901、2 号 101、3 号 101、3 号 301-501 邮编: 518129 t:(86-755) 253 sgs.china@sgs.com

Remark:

1. Since the data and / or information above division line of front page is provided by the applicant, the relevant results or conclusions of this report are only made for these data and / or information, SGS shall not be responsible for the authenticity and integrity of such data and information and the validity of the results and / or conclusions arising therefrom. Testing results only apply to the sample as received.
2. The below note(s) is(are) applicable in case the relevant superscript(s) and symbol(s) was(were) shown in the report:
N/A(NA) = Not Applicable N/C(NC) = Not Conducted N/T(NT) = Not Tested
3. Unless otherwise stated, the decision rule for conformity reporting is based on Binary Statement for Simple Acceptance Rule (w=0) stated in ILAC-G8:09/2019.

Sample Picture (As received)



SGS authenticate the photo on original report only

End of Report

